



## DECLARATION OF CONFORMITY

Philips Medizin Systeme  
Böblingen GmbH  
Hewlett-Packard Str. 2  
71034 Böblingen  
Germany

### Declares under our sole responsibility that the product:

Product Name: Battery for IntelliVue Patient Monitor  
Product Model Number or Designator: M4605A  
Product Options/Accessories: N/A

### To which this Declaration relates is in conformity with the provisions of Council Directive:

**2004/108/EC Electromagnetic compatibility (EMC)**

**and carries the CE-marking accordingly.**

#### Supplementary Information:

The products listed above have been tested in a typical configuration as described in the Manufacturer's accompanying documentation, and are compliant with the standards listed below and other related standards. Additionally the products listed above have been designed, manufactured, tested, and found to be compatible with the devices and accessories described by the manufacturer in the devices accompanying documentation:

EN 55011:2009 +A1:2010	Industrial, scientific and medical equipment - Radio-frequency disturbance characteristics – Limits and methods of measurement
EN 61000-4-2:2009	Electromagnetic compatibility (EMC) Part 4-2: Testing and measurement techniques – Electrostatic discharge immunity test
EN 61000-4-3:2006 +A1:2008+A2:2010	Electromagnetic compatibility (EMC) - Part 4-3: Testing and measurement techniques - Radiated, radio-frequency, electromagnetic field immunity test
EN 61000-4-6:2009	Electromagnetic compatibility (EMC) - Part 4-6: Testing and measurement techniques – Immunity to conducted disturbances, induced by radio-frequency fields

Signature:

Date: 30-March-2012

Printed Name: Hauke Schik  
Title: Director, Quality & Regulatory Affairs

Place of Issue: Böblingen